

CuroCell[®] M4 CX16

A dynamic mattress replacement system with good comfort and functionality



CuroCell® M4 CX16 with Top part Stone/bottom part CC

ALIN CER

| FACTS | OPTIONAL |
|--|---|
| Type: Dynamic mattress replacement system Height: 16 cm Recommended user weight: 0-200 kg Input voltage: 100-240 V / 50-60 Hz Sound level, control unit, max: 17 dBA⁽²⁾, 35 dBA⁽⁵⁾ | Durable bag Sheet holder |
| Warranty: 2 years PVC free | SIZES |
| Machine wash max 95 °C, tumble drying Cleaning of cover: Wipe with cleaning agent and/or disinfectants Chlorine resistant (1%) multi stretch cover (cover Stone with welded seams, 10%). | 80/85/90/100/105/120x200/210 x 16 cm |

CuroCell[®] M4 CX16

The CuroCell[®] M4 CX16 provides simplicity together with user comfort while aiding in the prevention and treatment of pressure ulcers/pressure injuries up to and including category 4, and unclassified pressure ulcers⁽¹⁾. It is designed for simplicity and provides features for good user comfort, reduction of shear forces and improvement of the microclimate.

The CuroCell[®] M4 CX16 mattress system includes a control unit with a user-friendly interface. The internal pressure of the mattress system automatically returns to an internal pressure preset by the product or user, regardless of user weight and position.

The control unit is equipped with the pulsating mode. With the pulsating mode, the benefits from constant low pressure (reduced maximum pressure) and the pressure-relieving effect of the dynamic mode are combined. Pulsating mode means that the pressure alternates and cells are emptied, but not completely. This gives a lower maximum pressure against the skin and slight movements in the mattress, which contributes to increased comfort and stability. The clinical effect of the pulsating mode has been investigated in scientific studies where both tissue perfusion, as well as prevention and therapeutic effect is measured⁽⁶⁾⁽⁷⁾. All mattresses are supplied with a four way stretch hygienic cover which meets standards for hygiene, quality and safety⁽³⁾⁽⁸⁾. All covers are manufactured from modern, recyclable materials and are vapour permeable⁽⁴⁾, which contributes to reduce the risk of skin maceration.

REFERENCES

(1) European Pressure Ulcer Advisory Panel, National Pressure Injury Advisory Panel and Pan Pacific Pressure Injury Alliance. Prevention and Treatment of Pressure Ulcers/Injuries: Clinical Practice Guideline. The International Guideline. Emily Haesler (Ed). EPUAP/NPIAP/PPPIA: 2019.

(2) EN ISO 11201 "Acoustics - Noise emitted by machinery and equipment - Determination of emission sound pressure levels", SP 2018.

(3) CDC (2003), Guidelines for Environmental Infection Control in Health-Care facilities, updated 2019.

(4) SS-EN ISO 15496:2004, DIN 53122-1.

(5) ISO 3746:2010 "Acoustics - Determination of sound power levels and sound energy levels of noise sources using sound pressure - Survey method using an enveloping measurement surface over a reflecting plane".

(6) Zwaenepoel, E., et al., Int Wound J, 2020 17(1).

(7) Worsley P, Dr Bostan L., Bader D. Investigation of the biomechanical and physiological responses to prolonged lying postures during three mattrees modes. Report from University of Southampton, 2018.

(8) SS 876 00 20:2017, Healthcare textiles - Mattresses - Specifications and requirements.



| FEATURES CONTROL UNIT | FEATURES MATTRESS |
|--|---|
| User friendly interface Audible and visual notifications Three program modes: pulsating, alternating and CLP (constant low pressure) Caring mode (max-firm) Pack&Go® function Silent running and vibration free Low voltage (12V power supply) Soft friction hangers with soft tip Air flow control; Constant temperature, good user comfort, energy reduction | Individually replaceable cells with safety mattress in HR-foam Heel function Multifunctional inner cover CPR-function with transport function Smart cable management Easy handling and usage Well protected zipper Always read the instruction for use prior to usage of the product. Care of Sweden and CuroCell are trademarks of |